

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

|  |                                      |
|--|--------------------------------------|
| DataMaster cdm S/N <u>130187</u>   |                                      |
| Supervisor/Operator Performing the Verification Procedure:                     |                                      |
| Name <u>DAN PASQUARIELLO</u>   | ID# <u>2620</u> Date <u>11-10-11</u> |
| A Agency <u>DILLINGHAM PD</u> Phone # <u>842-5354</u>                          |                                      |
| Instrument Location <u>DILLINGHAM PD</u>                                       |                                      |
| B Alco S/N <u>58428</u> Target Value <u>.080</u> High Pressure <u>1050 PSI</u> |                                      |
| Alco Test Values <u>.080</u> <u>.081</u><br>1st Alco 2nd Alco                  |                                      |
| Signature <u>[Signature]</u> <u>[Signature]</u><br>11/21/11                    |                                      |
| (OVER)   |                                      |

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 1<sup>st</sup> day of Feb, 2012 ak

[Signature] (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130187

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

NOVEMBER 10, 2011

OPERATOR'S NAME:

PASQUARIELLO, DAN

OPERATOR'S NUMBER: 2620

SUBJECT'S LAST NAME:

UOC

SUBJECT'S FIRST NAME/MI :

UOC

O.I.L. #: 1234567890

DEPT/AGENCY: DL61

CASE/REPORT: 1234567890

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 59428

### BREATH ANALYSIS

|                            |          |       |
|----------------------------|----------|-------|
| .000 ADJUSTED FOR 29.93 in |          |       |
| ALCO TARGET                | .000     | 10:32 |
| BLANK TEST                 | .000     | 10:33 |
| INTERNAL STANDARD          | VERIFIED | 10:33 |
| ALCO TV 29.93 in           | .000     | 10:33 |
| BLANK TEST                 | .000     | 10:34 |
| SUBJECT SAMPLE             | .000     | 10:36 |
| BLANK TEST                 | .000     | 10:37 |
| ALCO TV 29.93 in           | .001     | 10:37 |
| BLANK TEST                 | .000     | 10:38 |

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130187

NOVEMBER 10, 2011

TIME 10:39

### DIAGNOSTIC CHECK

|                  |          |
|------------------|----------|
| COMPUTER:        | OKAY     |
| PROGRAM:         | OKAY     |
| SOFTWARE DATE:   | 02/20/01 |
| HEATERS          |          |
| SAMPLE CHAMBER:  | 49c      |
| BREATH TUBE:     | 43c      |
| BAROMETER:       | 29.93 in |
| FLOW DETECTOR:   | OKAY     |
| PUMP             |          |
| HIGH SPEED:      | OKAY     |
| DETECTOR:        | OKAY     |
| FILTERS:         | OKAY     |
| QUARTZ STANDARD: | OKAY     |
| CALIBRATION:     | OKAY     |

### PRINTER TEST

!@#\$%^&\*+,-./0123456789:;<=>?@ABCODEFG  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~